


OPTIM[®] BLUE Quick Reference Guide

The product's label should be consulted before using this surface disinfectant.



OPTIM WIPES Ready to Use Wipes

- 1** Pull out OPTIM BLUE towelette at a 45 degree angle. Close lid.



- 2** If surface is not heavily soiled, clean and disinfect surface with one wipe and allow surface to remain wet for the contact time indicated on the label.



- 3** If surface is visibly soiled, clean surface with one towelette and discard. Use a second towelette to disinfect and allow surface to remain wet for the contact time indicated on the label.



OPTIM SOLUTION Liquid Spray

- 1** Spray OPTIM BLUE onto OPTIM dry wipe*.



- 2** If surface is not heavily soiled, clean and disinfect surface with one wipe and allow surface to remain wet for the contact time indicated on the label.



- 3** If surface is visibly soiled, clean surface with one towelette and discard. Use a second towelette to disinfect and allow surface to remain wet for the contact time indicated on the label.



*** IMPORTANT:**
Do not allow OPTIM BLUE to pool on surfaces. Gauze dipped in solutions for wiping may cause pooling and should not be used.



FOR BEST RESULTS:

- Do not over apply OPTIM BLUE.
- OPTIM BLUE may cause discolouration or damage if over applied on soft metals like brass or copper, anodized aluminum and soft plastic.
- Use of gloves with corn starch may cause yellowing of surfaces.
- Gloves with high levels of sulphur residue may cause a foul odour.
- Dental Association guidelines recommend the use of utility gloves when cleaning and disinfecting surfaces.

OPTIM BLUE is safe to use: No caution, danger or warning required on labels as it is the lowest toxicity in the category.

OPTIM BLUE has been tested in accordance with regulatory requirements as a One-Step Cleaner Disinfectant.

PLEASE REFER TO THE ORIGINAL CONTAINER FOR COMPLETE DIRECTIONS AND ADDITIONAL INFORMATION.

Procedure for disinfecting elastomeric impression materials with OPTIM BLUE Liquid

1. Once the impression is taken, rinse it well under running tap water to remove blood and saliva.
2. After rinsing, gently shake the impression within the sink basin to remove any adherent water with minimal splatter.
3. Spray the entire impression with OPTIM BLUE. Use the contact times as stated on the label.
4. Rinse the impression thoroughly with tap water to remove any excess disinfectant. After a thorough rinse, gently shake the impression within the sink basin to remove any adherent water with minimal splatter.
5. Once dry, the impression is ready for further processing.



Switching to OPTIM BLUE

- Give all surfaces a good rinse with soap and water before switching to OPTIM. After switching to OPTIM, surfaces need to be periodically rinsed with water (recommended no less than once per week)! Common practice for all chemical disinfectants is to periodically rinse surfaces to maintain the surface integrity.
- No mixing of chemistries.